

# Theranova

DESIGNED FOR: HDx MEMBRANE: MCO (PAES/PVP, BPA-free)

#### HDx THERAPY ENABLED BY THERANOVA\*

HDx therapy (expanded HD) is the next evolution in hemodialysis, as it targets the efficient removal of large middle molecules (25 kDa to < 60 kDa).<sup>1</sup> Indeed, many of them are linked to the development of inflammation, cardiovascular disease, and other co-morbidities in dialysis patients.<sup>2</sup> With HDx therapy, **Theranova** provides superior removal of large middle molecules compared with HD and HDF modalities and it does so using regular HD workflow and infrastructure.<sup>3</sup>

HDx therapy is enabled by the **Theranova** dialyzer series, which features an innovative membrane design that combines a permeability higher than that of regular high-flux dialyzers with effective selectivity for large proteins.<sup>4,5</sup>

#### PROVIDE EXPANDED HD, RETAIN HD SIMPLICITY

- Markedly greater clearances and intradialytic reduction ratios for middle molecules than regular HD at ordinary blood flow rates.<sup>3</sup>
- Superior removal of large middle molecules compared to HD and HDF modalities.<sup>3</sup>
- Limited albumin removal of between 1 and 4 grams per session.<sup>3</sup>
- Compatible with any HD monitor<sup>6,7</sup> and with standard dialysis.

# WITH BAXTER'S LATEST DIALYZER INNOVATION, COMING CLOSER TO THE NATURAL KIDNEY<sup>4,5</sup>

- High permeability to large middle molecules.
- Effective selectivity by size exclusion.
- Augmented internal filtration.
- Similar retention of endotoxins to other dialysis membranes of the same material.<sup>8</sup>

### CLINICAL EFFICIENCY AND PATIENT REPORTED OUTCOMES

- Pre-dialysis levels of b2 microglobulin and kappa and lambda free light chains were reduced at 3 months and this reduction was maintained up to 6 months evaluation period of HDx therapy in a multi-center observational study of 41 HD patients.<sup>9,‡</sup>
- Based on 2 large, long-term registry studies in prevalent HD patients there was an approximate 50% reduction in the number of patients meeting Restless Leg Syndrome (RLS) criteria after 6 months of HDx therapy<sup>11,12</sup> and an over 50% reduction in the number of patients meeting RLS criteria at 12 months.<sup>12</sup> A smaller before-after study found no difference in patient-reported symptom burden.<sup>10</sup>

#### \* Do not use Theranova dialyzers in HDF or HF mode

<sup>‡</sup>Based on data presented in a congress abstract – see reference for details.

## Based on data presented in a congress abstract - see reference for details. Restless leg syndrome was only one of several secondary endpoints.

## Theranova Specifications

MATERIALS	THERANOVA 400	THERANOVA 500
Membrane	Medium Cut Off	
	Polyarylethersulfone and Polyvinylpyrrolidone blend	
	BPA-free	
Potting	Polyurethane (PUR)	
Housing	Polycarbonate (PC)	
Gaskets	Silicone rubber (SIR)	
Protection caps	Polypropylene (PP)	
Sterilization	Steam (inside-out)	
Sterile barrier	Tyvek	
SPECIFICATIONS		
UF-Coefficient (mL/(h*mmHg))*	48	59
KoA urea*	1482	1630
Blood Compartment	91	105
volume (mL)	71	105
Minimum recommended	300	
priming volume (mL)	300	
Maximum TMP (mmHg)	600	
Recommended Q <sub>B</sub> (mL/min)	200-600	250-600
Storage conditions	<30°C (or <86°F)	
Units per box	24	
Gross/net weight (g)	229/170	246/190
MEMBRANE		
Effective Membrane Area (m²)	1.7	2.0
Fiber inner diameter (µm)	180	
Fiber wall thickness (µm)	35	
Sieving profile – before blood exposure <sup>4</sup>		
MWCO (cut-off) [kDa]	56 +/-3	
MWRO (rentation onset) [kDa]	9.4 +/- 0.2	
SIEVING COEFFICIENTS*		
Vitamin B12 (1,4 kDa)	1.0	
Inulin (5,2 kDa)	1.0	
B <sub>2</sub> -microglobulin (11,8 kDa)	1.0	
Myoglobin (17 kDa)	0.9	
	0.000	

CLEARANCES IN VITRO (mL/min)*	THERANOVA 400	THERANOVA 500
<b>Urea (60 Da)</b> (Q <sub>B</sub> -Q <sub>D</sub> , mL/min)		
200/500	198	199
300/500	282	285
400/500	344	351
400/800	376	381
500/800	445	454
Phosphate (95 Da)		
200/500	192	194
300/500	261	267
400/500	311	320
400/800	345	354
500/800	400	413
Creatinine (113 Da)		
200/500	194	196
300/500	269	274
400/500	323	331
400/800	357	365
500/800	416	428
Vitamin B12 (1.4 kDa)		
200/500	164	169
300/500	207	215
400/500	239	249
400/800	267	280
500/800	301	317
Inulin (5.2 kDa)		
200/500	133	139
300/500	161	170
400/500	183	193
400/800	204	216
500/800	225	241
Cytochrome C (12 kDa)		
200/500	122	128
300/500	146	155
400/500	165	175
400/800	183	196
500/800	202	217
Myoglobin (17 kDa)		
200/500	104	110
300/500	123	130
400/500	137	147
400/800	152	163
500/800	166	180

\* According to EN 1283/ISO 8637:

Albumin (66,4 kDa)

- UF-Coefficient: measured with bovine blood, Hct 32%, Pct 60g/L, 37°C

– KoA urea: calculated at  $Q_B$ =300 mL/min,  $Q_D$ =500mL/min, UF=0 mL/min

– Sieving coefficients: measured with human plasma,  $Q_B$ =300 mL/min, UF=60 mL/min

- Clearances In-Vitro: measured at UF=0 mL/min, ±10% (±20% Cyt. C, ±30% Myo.)

1. Ronco C, et al. The rise of Expanded Hemodialysis. Blood Purif 2017; 44:I-VIII.

2. Hutchison CA, et al. The Rationale for Expanded Hemodialysis Therapy (HDx). Contrib Nephrol 2017; 191:142-52.

3. Kirsch AH, et al. Performance of hemodialysis with novel medium cut-off dialyzers. Nephrol Dial Transpl 2017; 32(1):165-72.

4. Boschetti-de-Fierro A, et al. MCO membranes: Enhanced Selectivity in High-Flux Class. Scientific Reports 2015; 5:18448.

5. Zweigart C, et al. Medium cut-off membranes - closer to the natural kidney removal function. Int J Artif Organs 2017; 40(7):328-334.

0.008

6. Baxter. Data on file. Theranova Limited Controlled Distribution Report. 2016.

7. Baxter. Theranova 400/500 Instructions For Use. N50 648 rev 003, 2017-05-29.

8. Schepers E, Glorieux G, Eloot S, et al. Assessment of the association between increasing membrane pore size and endotoxin permeability using a novel experimental dialysis simulation set-up. BMC Nephrology. 2018; 19:1.

Cantaluppi V et al. Removal of large-middle molecules on expanded hemodialysis (HDx): a multicentric observational study of 6 months follow-up. ASN 2018 Kidney Week Abstract TH-P0357.
Krishnasamy R et al. Trial evaluating mid cut-off value membrane clearance of albumin and light chains in hemodialysis patients (REMOVAL-HD): a safety and efficacy study. ASN 2018 Kidney Week Abstract TH-P0353.

11. Sanabria M et al. Quality of life reported by patients with expanded hemodialysis by the Theranova dialyzer in RTS Colombia. ASN 2018 Kidney Week Abstract TH-P0296.

12. Sanabria M, et al. Patient-Reported Outcome Measures (PROMs) and Expanded Hemodialysis (HDx) with Medium Cut-Off Dialyzers in a Large Cohort of Patients in Colombia: The COREXH Study. ASN 2018 Kidney Week Abstract FR-P0493.

#### For safe and proper use of the device, please refer to the Instructions for Use.

#### **Baxter Corporation**

7125 Mississauga Road, Mississauga, ON L5N 0C2 Canada 1-888-719-9955 Visit us at www.baxter.ca

Baxter, MCO and Theranova are trademarks
of Baxter International Inc. or its subsidiaries.